

Philips Medical Systems

October 3, 2005

Rita Chew RDMS,RDCS,RVT
President, Essoula LLC
4917 Evergreen Way #5
Everett WA 98203
425-308-4379

Ref: Essoula Ultrasound Gel Compatibility

Dear Rita,

Essoula Comfort, Aromatherapy Enhance Ultrasound Gel, was evaluated for compatibility with Philips designed ultrasound probes. The results of that testing confirmed that the product is compatible with the following probes.

D2cwc	L7-4
D2tcd	L12-5 50mm
D5cwc	L17-5
3D6-2	L12-3
3D8-4	L15-7io
3D9-3v	PA4-2
C8-4v	S4-1 (P4-1)
C9-5(ec) ICT	S12-4
C5-2	S8-3
C8-5	S4-2
C9-4	S3-1
L8-4	PA4-2
L9-3	

NOTE: This product is not approved for use on TEE probes.

Sincerely,



Kevin G. Wickline
Chemist - R&D Engineering
Philips Ultrasound
1 Echo Drive
Reedsville, PA 17084
Phone: 717-667-5056
Fax: 717-667-5169
email: kevin.wickline@philips.com

